

Model 7500 Series



Easy to use. Economical. Reliable performance.

Nonin's 7500 pulse oximeter has a unique low-profile design that enables it to fit easily on a tabletop or into its own carrying case. The stylish, compact 7500's simple functionality, minimum 16-hour battery life, 4-hour quick recharge, and minimum of 70 hours of memory makes it well-suited for a variety of clinical and homecare applications. Like all Nonin portable oximeters, the 7500 provides a wide range of display indicators that provide reliable readings.

The 7500's flexible alarm management options include patient security mode and a user-defined default that offers customizable patient settings. The 7500's analog output capability allows it to interface easily with other equipment.

A variety of single-use and reusable sensors are available for use with the Model 7500. Nonin Medical's PureLight® sensors produce accurate and dependable SpO₂ readings from patient-to-patient and sensor-to-sensor. Use only Nonin branded sensors.

Model 7500 Series

Product Highlights

- Convenient Lightweight, compact
- Easy-To-Use Bright LED displays and large display screen offer clear readability
- **Durable** Rugged design
- **Powerful** Over 70 hours of memory available for extended monitoring
- Versatile Monitors neonate to adult patients in acute care to homecare environments
- **Flexible** Use as a stand alone device or interface with other equipment

Magnetic Resonance (MR) Model 7500 FO

The 7500FO (fiberoptic) pulse oximeter was specifically developed for MR environments. Use in fixed-site or mobile MR facilities for monitoring infant, pediatric and adult patients.



User-Friendly Interface

The 7500 provides a variety of display indicators that ensure reliable readings.



Pulse Strength Bar Graph – Color coded bar graph assists in assessing the patient's status. An audible tone sounds with each pulse. The pitch changes with the patient's saturation level.



Pulse Quality Indicator – Detects patient movement or signal quality changes that may require sensor site stabilization.



Sensor Fault Indicator – Indicates when a sensor has become disconnected, misaligned, failed, or is incompatible with the monitor.

Specifications

O	ian Dissiles 5		0.1000/.5.0	
Oxygen Saturati		•		
Pulse Rate Rang	je		18–321 BPIVI	
Displays Dulce Strongt	.h		LED Bar graph tri s	alar sagmants
Pulse Strength				
Alarm Indicator				
Alarm Silenced			,	
Numeric Displays				
Low Battery Saturation Accuracy (A _{rms}) ^a				
Saturation Accu	iracy (A _{rms})			Manuatas
No Madian	Danisalala	Financija	Adults/Pediatrics	
No Motion:	Reusable	Finger Clip	•	± 3 digits
Madian		6000 Series		± 3 digits
Motion:		Finger Clip	± 2 digits	NA
Pulse Rate Accu	-		10 200 DDM	
No Motion:				
Motion:				
Low Perfusion:			40–240 BPM Adults/Pediatrics Neonates	
	5 11	F: CI:		
No Motion:		Finger Clip	-	± 3 digits
	'	6000 Series	9	± 3 digits
Motion:		Finger Clip		± 5 digits
Low Perfusion:		Finger Clip		± 3 digits
		6000 Series	•	± 3 digits
Memory				
Temperature (Operating)				
(Storage/Transportation)				
Humidity (Operating)			•	
(Storage/Transportation)				
Altitude (Operating)			·	
Hyperbaric Pressure			· ·	
•	nents (Mains)		100–240 VAC 50–6	0 Hz
Internal Power				
Battery			, ,	
Operating Life (fully charged battery)				
Storage Life			•	
Recharge Rate				
Dimensions				
			92 mm (3.6") H x 142 mm (5.6") D	
Weight				
Warranty			3 years	
Classification pe	er IEC 60601	-1/CSA601.1/U	L60601-1	
Type of Protection			Class II (when on AC power with MPP30 battery	
			charger), Internally powered (on battery power)	
Degree of Prote	ction		Type BF-Applied Par	t
Mode of Operat	tion		Continuous	
Enclosure Degre	ee of Ingress	Protection	IPX2	
Analog Outputs				
SpO ₂ Output Range			.0-1 VDC (0–100% SpO $_2$), 1.27 VDC (out of track)	
Pulse Rate Output Range			0-1 VDC (0–300 BPM), 1.27 VDC (out of track)	
Event Marker				
Accuracy			± 2% (SpO ₂), ± 5% (Pulse Rate)	
Load Current			. 2 mA maximum	

Specifications are subject to change without notice.

 $^{\text{a}}$ \pm 1 A_{rms} represents approximately 68% of measurements

Quality systems certified to ISO 13485, and CE per Annex II of EC Directive No. 93/42/EEC as amended by 2007/47/FC.

Authorized Distributor:



PureSAT' Sp





Prins Hendriklaan 26 1075 BD Amsterdam • Netherlands

Tel: +31 (0)13-79 99 040 Fax: +31 (0)13-79 99 042 E-mail: infointl@nonin.com



Nonin Medical, Inc.

Fax: +1.763.577.5521

E-mail: info@nonin.com

13700 1st Avenue North

Plymouth, MN • 55441-5443 • U.S.A.

Tel: +1.763.553.9968 1.800.356.8874

